

APPENDIX 1F.2: EQUIPMENT LIST – CATEGORY 1, 2, 3

NOTE: THE OBLIGATIONS AND RESPONSIBILITIES OF THE PARTIES RELATED TO CATEGORY 1, 2, AND 3 EQUIPMENT ARE SET OUT IN APPENDIX 1F.1: EQUIPMENT RESPONSIBILITY.

APPENDIX 1F.2 EQUIPMENT
 MECHANICAL DETAILS CAT. 1, 2, 3 - BLH
 BASIS OF DESIGN ONLY

								FOOTPRINT mm				MOUNTING					ELECTRICAL - DATA INFO TO BE CONFIRMED WITH NHA IT				PLUMBING				MEDICAL GASES				VENTING	ALARM									
ROOMDESC	ROOM NO	ITEM ID	QTY	ITEM NAME	DESCRIPTION	DESC2	CAT #	CUT SHEET INFO.	H	W	D	WGT	MOBILE	BENCH	WALL	REC'D	CEIL	MOUNT NOTES	V	PH	AMP	FREQ.	DIR CIR	EMER PWR	ELECT NOTES	HOT WATER	COLD WATER	DRAIN SIZE	WATER NOTES	O2	N2O	AIR	VAC	MEDICAL GAS NOTES	VENT NOTES	LOC.	OTHER NOTES		
Laboratory	219	100465	1	CABINET	BIOLOGICAL SAFETY, W/STAND	CLASS II TYPE B2	2	Basis of Design cut sheet	2279	2030	815	438		XX					115	1	12.4/20	60														TBD If unit is A2 recirculating or B2 vended out		TBD if unit is A2 recirculating or B2 vented out	
Laboratory	219	100350	4	CAN	GARBAGE		1	Basis of Design cut sheet	508	279	387	7.26		XX										N/A															
Laboratory	219	100041	2	CART	UTILITY, 3 SHELF	27 X 16 X 32"H	1	Basis of Design cut sheet	817	699	413	13.2		XX										N/A															
Laboratory	219	100486	2	CENTRIFUGE	BENCHTOP	TO BE DETERMINED	1	Basis of Design cut sheet	418	457	750	90							120	1	15	60																	
Laboratory	219	100353A	2	DISPENSER	GLOVE BOX	COUNTERTOP	1	See cut sheet 100353	254	457	102			XX			Dispenser will sit on counter, does not need to be wall mounted							N/A															
Laboratory	219	100985	5	HOOK			3	Cut Sheet Not Available							XX									N/A															
Laboratory	219	100519	16	LAB BENCHES			3	Cut Sheet Not Available									To be determined, this may be millwork	115	1		60																		
Laboratory	219	100487	2	MICROSCOPE			1	Cut Sheet Not Available	378	235	500	7.3			XX				120	1	1.5	60																	
Laboratory	219	100343	1	MOUNT	WALL, TILT	TELEVISION up to 46"	3	Basis of Design cut sheet							XX									N/A															
Laboratory	219	100083	1	READER	CARDIAC		1	Cut Sheet Not Available							XX																								
Laboratory	219	100877	1	REFRIGERATOR	BLOOD 1/2 SIZE	ALARMED	1	Basis of Design Cut Sheet	876	610	613				XX		recommend 3" minimum be left on all sides for air flow	120	1	3.2/15	60																	Alarmed to Plant Services	
Laboratory	219	100869	1	STAINER	PROGRAMMABLE, MULTIPURPOSE		1	Cut Sheet Not Available	250	570	540	16.8			XX				110-240	1	0.8	60		XX															
Laboratory	219	100473	3	UPS	LAB		1	Cut Sheet Not Available	185	115	213	5.7					Sits on floor	120/240	1		60		XX															Future data drop	

APPENDIX 1F.2 EQUIPMENT
 MECHANICAL DETAILS CAT. 1, 2, 3 - BLH
 BASIS OF DESIGN ONLY

								FOOTPRINT mm				MOUNTING					ELECTRICAL - DATA INFO TO BE CONFIRMED WITH NHA IT						PLUMBING				MEDICAL GASES				VENTING	ALARM									
ROOMDESC	ROOM NO	ITEM ID	QTY	ITEM NAME	DESCRIPTION	DESC2	CAT #	CUT SHEET INFO.	H	W	D	WGT	MOBILE	BENCH	WALL	REC'D	CEIL	MOUNT NOTES	V	PH	AMP	FRO.	DIR CIR	EMER PWR	ELECT NOTES	HOT WATER	COLD WATER	DRAIN SIZE	WATER NOTES	O2	N2O	AIR	VAC	MEDICAL GAS NOTES	VENT NOTES	LOC.	OTHER NOTES				
Clean Holding	246	100089	2	CART	SUPPLY	5'	1	Basis of Design cut sheet	1676	1676	660	93	XX											N/A																	
Clean Holding	246	100388	1	COMMODE	/SHOWER	(standard / recliner to be determined)	1	Basis of Design cut sheet	1016	737	635	18.1	XX					Clears 16 3/4" high toilet, 34" L with footrest down.						N/A																	
Clean Holding	246	100217	1	RACK	BEDPAN LINER DISPENSER		2	Basis of Design cut sheet	991	432	330	10.9			XX			be sure bottom of rack is above cart height						N/A																	
Clean Holding	246	100217A	1	RACK	BEDPAN LINER MALE URINAL		2	Basis of Design cut sheet	965	279	381	10.4			XX			be sure bottom of rack is above cart height						N/A																	
Treatment bay	247	100767	1	ACCESSORIE S	FOR MOUNTING	BASKETS/SHELVES/ADAPTERS	2	Basis of Design Cut Sheet	931	130	100		XX					Mounted onto headwall						N/A																	
Treatment bay	247	100350	1	CAN	GARBAGE		1	Basis of Design cut sheet	508	279	387	7.26	XX											N/A																	
Treatment bay	247	100233	1	CART	PROCEDURE		1	Basis of Design cut sheet	1029	863	559		XX											N/A																	
Treatment bay	247	100098	1	CONTAINER	SHARPS	CHILD PROOF	2	Basis of Design cut sheet	323	267	184			XX	XX			Wall or headwall mounted						N/A																	
Treatment bay	247	100353	1	DISPENSER	GLOVE BOX	TRIPLE	2	Basis of Design cut sheet	254	457	102			XX	XX			size approximate - to be confirmed by NHA						N/A																	
Treatment bay	247	100347	1	DISPENSER	PAPER TOWEL	HANDS FREE	2	Basis of Design cut sheet	410	321	259				XX			Suggested mounting height at 52"	115	1		60		may be battery operated																	
Treatment bay	247	100348	1	DISPENSER	SOAP	HANDS FREE	2	Cut Sheet Not Available	356	127	127				XX			size approximate - to be confirmed by NHA						N/A																	
Treatment bay	247	100030	1	FLOWMETER	AIR		1	Basis of Design cut sheet	178	50.8	102				XX									N/A													XX				
Treatment bay	247	100020	1	FLOWMETER	OXYGEN		1	Basis of Design cut sheet	178	50.8	102				XX									N/A												XX					
Treatment bay	247	100180	1	HAND SANITIZER	WATERLESS		2	Basis of Design cut sheet	233	132	115				XX									N/A																	
Treatment bay	247	100498	1	HEADWALL	W/MOUNTING RAIL		3	Basis of Design cut sheet							XX			Size varies with room/bed layout	115	1		60	XX													XX	XX	XX			

								FOOTPRINT mm				MOUNTING					ELECTRICAL - DATA INFO TO BE CONFIRMED WITH NHA IT						PLUMBING				MEDICAL GASES				VENTING	ALARM									
ROOMDESC	ROOM NO	ITEM ID	QTY	ITEM NAME	DESCRIPTION	DESC2	CAT #	CUT SHEET INFO.	H	W	D	WGT	MOBILE	BENCH	WALL	REC'D	CEIL	MOUNT NOTES	V	PH	AMP	FREQ.	DIR CIR	EMER PWR	ELECT NOTES	HOT WATER	COLD WATER	DRAIN SIZE	WATER NOTES	O2	N2O	AIR	VAC	MEDICAL GAS NOTES	VENT NOTES	LOC.	OTHER NOTES				
Patient room (single occupancy)	256	100350	1	CAN	GARBAGE		1	Basis of Design cut sheet	508	279	387	7.26	XX												N/A																
Patient room (single occupancy)	256	100051	1	CLOCK	WALL	W/SECOND HAND	3	Cut Sheet Not Available							XX										Battery Operated																
Patient room (single occupancy)	256	100098	1	CONTAINER	SHARPS	CHILD PROOF	2	Basis of Design cut sheet	323	267	184			XX	XX			Wall or headwall mounted							N/A																
Patient room (single occupancy)	256	100353	1	DISPENSER	GLOVE BOX	TRIPLE	2	Basis of Design cut sheet	254	457	102			XX	XX			size approximate - to be confirmed by NHA							N/A																
Patient room (single occupancy)	256	100030	1	FLOWMETER	AIR		1	Basis of Design cut sheet	178	50.8	102				XX										N/A								XX								
Patient room (single occupancy)	256	100020	2	FLOWMETER	OXYGEN		1	Basis of Design cut sheet	178	50.8	102				XX										N/A								XX								
Patient room (single occupancy)	256	100791	1	HAMPER	LINEN	SINGLE	1	Basis of Design Cut Sheet	838	457	457			XX											N/A																
Patient room (single occupancy)	256	100180	1	HAND SANITIZER	WATERLESS		2	Basis of Design cut sheet	233	132	115				XX										N/A																
Patient room (single occupancy)	256	100214	1	LIFT	PATIENT	CEILING MOUNT	1	Cut sheet provided																																	
Patient room (single occupancy)	256	100032	1	MOUNT	MONITOR		1	Basis of Design cut sheet	406	88.9					XX			Final quantity of monitors to be used with mounts to take place with Biomed. Do not purchase until review is complete	120	1	1	60		XX																	
Patient room (single occupancy)	256	100343	1	MOUNT	WALL, TILT	TELEVISION up to 46"	3	Basis of Design cut sheet							XX										N/A																
Patient room (single occupancy)	256	100255	1	PILLOW SPEAKER			1	Cut Sheet Not Available						XX	XX							60																			
Patient room (single occupancy)	256	100215A	1	RAIL	GANTRY	OVERHEAD LIFT	3	See cut sheet 100215									XX								N/A																
Patient room (single occupancy)	256	100028	2	REGULATOR	SUCTION		1	Basis of Design cut sheet	178	50.8	102				XX										N/A											XX					

